

Dosimetry Characterization of a Brachytherapy Seed by Thermoluminescence Dosimetry in Liquid Water

R. C. Taylor, G. S. Ibbott, S. Lampe, W. Bivens, and N. B. Tolani

University of Texas MD Anderson Cancer Center



Purpose:

A primary goal of this project was to resolve discrepancies in the reported values of the dose rate constant for ^{137}Cs (model Cs-1), and to use liquid water as a phantom to measure a complete set of dosimetric parameters. This method avoids the uncertainties introduced by use of water-equivalent plastics.

Introduction:

The updated AAPM Task Group 43 Report (TG-43U1)¹ for brachytherapy dosimetry recommends the use of dosimetric parameters measured in, or related to, water. Measurements of dosimetric parameters are often made using thermoluminescence dosimetry (TLD). However, TLD measurements in water are quite challenging to perform for a number of reasons. As such, brachytherapy sources are generally characterized in water-equivalent plastics such as Solid Water™ (SW). The dose measured in SW must then be converted to dose in liquid water using a correction factor based on Monte Carlo (MC) calculations. Errors can be introduced by this procedure if the chemical composition of the SW material used for measurement is not identical to that used for MC simulation.²

To avoid the additional sources of uncertainty associated with measurements in water-equivalent materials, a set of plastic jigs was designed to support TLD capsules around a brachytherapy source in liquid water.³ The jigs were used to characterize a new model of brachytherapy source, the ^{137}Cs source (model Cs-1 Rev. 2) produced and marketed by IsoRay Medical, Inc. (Richland, Washington).

^{137}Cs produces a mean photon energy of ~ 30.4 keV, and is of interest because it exhibits desirably shorter $T_{1/2}$ of 9.7 days⁴ compared to 60 days for ^{125}I and 17 days for ^{109}Pd .⁵ A radiobiological model suggests advantages of short $T_{1/2}$ in permanent brachytherapy implants.⁶

Several reports of the dosimetric characteristics of ^{137}Cs have been published.⁷⁻¹⁰ There are significant differences among the values for dose rate constant reported previously.⁷⁻¹⁰ Our results measured in liquid water are compared against those from previous publications.⁷⁻¹⁰

Methods & Materials:

^{137}Cs source: Fig.1 shows a cut away view of the seed (model Cs-1 Rev.2) through its longitudinal axis. The source has been designed with outer dimensions similar to those of existing ^{125}I and ^{109}Pd seeds. The radioactive core is made of a gold wire surrounded by an inorganic substrate to which ^{137}Cs is chemically bound.

A total of 13 seeds (^{137}Cs) were used in this study. Air-kerna strength of each source was determined by the M.D. Anderson ADCL. The sources used in this study had initial strengths of 11-15 U.




Fig. 1: Schematic of the cross-section through the axis of ^{137}Cs model CS-1.

TLD capsules: Measurements were performed using TLD capsules. A capsule (Fig.2) consists of a glass capillary tube containing ~ 13 mg of lithium fluoride powder (TLD-100). The powder volume corresponds to 7.0 mm in length and 1.4 mm in diameter.




Fig. 2: Picture and schematic of TLD capsule.

TLD reading: A manual TLD reader (Harshaw model 3500) and standard procedures^{11,12} developed by the Radiological Physics Center (RPC) were employed to read irradiated TLD. The TLD signal (TLD reading/mass) was converted to dose following the RPC convention.^{11,12} Dose response of TLD was calibrated in a ^{60}Co beam, whose output was measured using the TG51 protocol.¹³ Relative energy response of TLD for 28-30 keV versus ^{60}Co photons was previously determined to be $0.707 \pm 2.5\%$.¹⁴ Other corrections due to the capsule's glass wall, and TLD volume's finite size were also incorporated.¹⁵

Dose-rate constant: Measurements were made employing a specially designed jig.³ This jig holds 24 TLD capsules equally spaced in a circle of 1.00 cm radius around the seed. During irradiation, the seed mounted coaxially on tip of a graphite rod was vertical (parallel to all TLD) and at the exact center of TLD array. For irradiation, the jig supporting the source and TLD capsules was submerged in water so that the source and capsules were surrounded by at least 10 cm of water in all directions.¹⁴ After irradiation, the TLD were read, and the TLD signal was converted to dose. A correction for decay during the irradiation was applied. The dose rate constant was determined from an average of the 24 TLD signals in each irradiation run (total 8 runs for water and SW). The results are shown in Table-1.

Anisotropy Function: The same jig, as used for measurement of Λ , was employed. However, the seed was glued perpendicularly to the tip of the graphite support rod. The jig has 24 equally spaced TLD positioning holes along each of the 5 concentric circles (radii 1-7 cm). A set of 24 TLD was positioned along one circle at a time to prevent TLD's mutual shadowing.

Measurements were repeated with 4 different seeds. In each irradiation run, symmetry of TLD in 4 quadrants provided 4 data points per angle (θ). The active length (L) of 4.0 mm, provided by the manufacturer, was employed in determination of the geometry factor $G(r, \theta)$.

Radial Dose Function: Another jig³, designed for this measurement, holds TLD capsules in 3-armed spiral pattern to avoid TLD's mutual shadowing. The jig holds the seed (^{137}Cs) to be irradiated, parallel to the TLD capsules. Each spiral arm holds 15 TLD capsules positioned at distances 0.5-10 cm from the source axis. Following irradiation, each TLD measurement was converted to dose as explained previously.

Results & Conclusions:

Dose rate Constant: Our results from 8 irradiation runs (2 per seed) are presented in Table-1, demonstrating an average dose rate constant of $1.078 \pm 3.3\%$. Comparison with the values reported by other investigators is also provided.

Source ID	A	Source A	Reference	Comments
PS02-01-01	1.074	1.081	1.081	Used as water design of the source.
PS02-01-01	1.074	1.081	1.081	Estimated by geometry correction.
PS02-01-01	1.074	1.081	1.081	TLD dosimetry in Solid Water™.
PS02-01-01	1.074	1.081	1.081	MC calculations.
CS000401	1.087	1.087	1.087	MC calculations.

Table-1: Values of dose rate constant.

Anisotropy Function: Table-2 summarizes the measured values of $F(r, \theta)$ and compares them with the values reported by Rivard.¹⁰ The 1D anisotropy function was determined by simple numerical integration from the 2D dose-rate measurements.

r (cm)	1 cm		2 cm		3 cm		4 cm		5 cm		7 cm	
	MDACC	Report Rate										
0	1.000	1.000	1.000	1.000	1.000	1.000	1.000	1.000	1.000	1.000	1.000	1.000
10	0.976	0.976	0.976	0.976	0.976	0.976	0.976	0.976	0.976	0.976	0.976	0.976
20	0.952	0.952	0.952	0.952	0.952	0.952	0.952	0.952	0.952	0.952	0.952	0.952
30	0.928	0.928	0.928	0.928	0.928	0.928	0.928	0.928	0.928	0.928	0.928	0.928
40	0.904	0.904	0.904	0.904	0.904	0.904	0.904	0.904	0.904	0.904	0.904	0.904
50	0.880	0.880	0.880	0.880	0.880	0.880	0.880	0.880	0.880	0.880	0.880	0.880
60	0.856	0.856	0.856	0.856	0.856	0.856	0.856	0.856	0.856	0.856	0.856	0.856
70	0.832	0.832	0.832	0.832	0.832	0.832	0.832	0.832	0.832	0.832	0.832	0.832
80	0.808	0.808	0.808	0.808	0.808	0.808	0.808	0.808	0.808	0.808	0.808	0.808
90	0.784	0.784	0.784	0.784	0.784	0.784	0.784	0.784	0.784	0.784	0.784	0.784
100	0.760	0.760	0.760	0.760	0.760	0.760	0.760	0.760	0.760	0.760	0.760	0.760

Table-2: Values of $F(r, \theta)$ measured in liquid water.

Our measured values of $F(r, \theta)$ are graphically presented below in Fig.3 along with the values reported by Murphy et. al.⁷ and Rivard.¹⁰ Our values agree better with Rivard's than for Murphy's values.








Fig. 3: Values of $F(r, \theta)$ measured in liquid water.

Radial Function: Our measured values of $g(r)$ based on line-source approximation, along with the values reported by Murphy et. al.⁷ and Rivard¹⁰ for comparison, are shown in Figure 4. Generally good agreement is seen over the range in which measured data were reported by Murphy.⁷ Better agreement is seen with the MC calculations reported by Rivard.¹⁰




Fig. 4: Radial dose function measured in water.

Table-3 presents our measured values of $g(r)$. Comparison with the values reported by Murphy et. al.⁷ and Rivard¹⁰ is also shown.

r (cm)	MDACC Mean in liquid water	Rivard MC in liquid water	Murphy Mean in SW	Murphy MC in liquid water
0.000	1.000	1.000	1.000	1.000
0.875	0.985	0.985	0.985	0.985
1.750	0.970	0.970	0.970	0.970
2.625	0.955	0.955	0.955	0.955
3.500	0.940	0.940	0.940	0.940
4.375	0.925	0.925	0.925	0.925
5.250	0.910	0.910	0.910	0.910
6.125	0.895	0.895	0.895	0.895
7.000	0.880	0.880	0.880	0.880
7.875	0.865	0.865	0.865	0.865
8.750	0.850	0.850	0.850	0.850
9.625	0.835	0.835	0.835	0.835
10.500	0.820	0.820	0.820	0.820
11.375	0.805	0.805	0.805	0.805
12.250	0.790	0.790	0.790	0.790
13.125	0.775	0.775	0.775	0.775
14.000	0.760	0.760	0.760	0.760
14.875	0.745	0.745	0.745	0.745
15.750	0.730	0.730	0.730	0.730
16.625	0.715	0.715	0.715	0.715
17.500	0.700	0.700	0.700	0.700
18.375	0.685	0.685	0.685	0.685
19.250	0.670	0.670	0.670	0.670
20.125	0.655	0.655	0.655	0.655
21.000	0.640	0.640	0.640	0.640
21.875	0.625	0.625	0.625	0.625
22.750	0.610	0.610	0.610	0.610
23.625	0.595	0.595	0.595	0.595
24.500	0.580	0.580	0.580	0.580
25.375	0.565	0.565	0.565	0.565
26.250	0.550	0.550	0.550	0.550
27.125	0.535	0.535	0.535	0.535
28.000	0.520	0.520	0.520	0.520
28.875	0.505	0.505	0.505	0.505
29.750	0.490	0.490	0.490	0.490
30.625	0.475	0.475	0.475	0.475
31.500	0.460	0.460	0.460	0.460
32.375	0.445	0.445	0.445	0.445
33.250	0.430	0.430	0.430	0.430
34.125	0.415	0.415	0.415	0.415
35.000	0.400	0.400	0.400	0.400
35.875	0.385	0.385	0.385	0.385
36.750	0.370	0.370	0.370	0.370
37.625	0.355	0.355	0.355	0.355
38.500	0.340	0.340	0.340	0.340
39.375	0.325	0.325	0.325	0.325
40.250	0.310	0.310	0.310	0.310
41.125	0.295	0.295	0.295	0.295
42.000	0.280	0.280	0.280	0.280
42.875	0.265	0.265	0.265	0.265
43.750	0.250	0.250	0.250	0.250
44.625	0.235	0.235	0.235	0.235
45.500	0.220	0.220	0.220	0.220
46.375	0.205	0.205	0.205	0.205
47.250	0.190	0.190	0.190	0.190
48.125	0.175	0.175	0.175	0.175
49.000	0.160	0.160	0.160	0.160
49.875	0.145	0.145	0.145	0.145
50.750	0.130	0.130	0.130	0.130
51.625	0.115	0.115	0.115	0.115
52.500	0.100	0.100	0.100	0.100
53.375	0.085	0.085	0.085	0.085
54.250	0.070	0.070	0.070	0.070
55.125	0.055	0.055	0.055	0.055
56.000	0.040	0.040	0.040	0.040
56.875	0.025	0.025	0.025	0.025
57.750	0.010	0.010	0.010	0.010
58.625	0.000	0.000	0.000	0.000
59.500	0.000	0.000	0.000	0.000
60.375	0.000	0.000	0.000	0.000
61.250	0.000	0.000	0.000	0.000
62.125	0.000	0.000	0.000	0.000
63.000	0.000	0.000	0.000	0.000
63.875	0.000	0.000	0.000	0.000
64.750	0.000	0.000	0.000	0.000
65.625	0.000	0.000	0.000	0.000
66.500	0.000	0.000	0.000	0.000
67.375	0.000	0.000	0.000	0.000
68.250	0.000	0.000	0.000	0.000
69.125	0.000	0.000	0.000	0.000
70.000	0.000	0.000	0.000	0.000
70.875	0.000	0.000	0.000	0.000
71.750	0.000	0.000	0.000	0.000
72.625	0.000	0.000	0.000	0.000
73.500	0.000	0.000	0.000	0.000
74.375	0.000	0.000	0.000	0.000
75.250	0.000	0.000	0.000	0.000
76.125	0.000	0.000	0.000	0.000
77.000	0.000	0.000	0.000	0.000
77.875	0.000	0.000	0.000	0.000
78.750	0.000	0.000	0.000	0.000
79.625	0.000	0.000	0.000	0.000
80.500	0.000	0.000	0.000	0.000
81.375	0.000</			