

TG-51 Worksheet A: Photon Beams

1. Site data

Institution: My Cancer Center
 Physicist: Smart Physicist
 Date: July 2000
 Accel or ^{60}Co Mfr.: Varian
 Model & serial number: Clinac 2300
 Nominal photon energy/beam identifier: 23 MV

2. Instrumentation

a. Chamber model: NEL 2571
 Serial number: # 2952
 Cavity inner radius (r_{cav} , Table III): 0.315 cm

Waterproof: yes no
 If no, is waterproofing ≤ 1 mm PMMA or thin latex?: yes no

b. Electrometer model: Kethley 602
 Serial number: 30102A
 i. P_{elec} , electrom. Corr factor (Sec. VII.B): 0.998×10^{-8} C/C or C/rdg.
 c. Calibration factor $N_{D,w}^{60\text{Co}}$ (Sec.V): 4.533×10^7 Gy/C (or Gy/rdg)
 Date of report (not to exceed 2 years): 11/11/99

3. Measurement Conditions (10 x 10 cm², point of measurement at 10 cm depth (water equivalent))

a. Distance (SSD or SAD): 100 cm SAD or SSD
 b. Field size: 10 x 10 cm²
 on surface (SSD setup)
 at detector (SAD setup):
 c. Number of monitor units: 200 MU (min for ^{60}Co)

4. Beam Quality (Sec. VIII.B – not needed for ^{60}Co)

If energy <10 MV, use no lead foil.

Measure % $dd(10)$ [% depth-dose at 10 cm depth for curve shifted upstream by $0.6r_{\text{cav}}$]

Field size 10 x 10 cm² on surface, SSD = 100 cm: yes no

a. % $dd(10)_x = \% dd(10)$

If energy ≥ 10 MV

Distance of 1 mm lead foil phantom surface 50 \pm 5 cm 30 \pm 1 cm

Measure % $dd(10)_{\text{Pb}}$ [% depth-dose at 10 cm depth for curve shifted upstream by $0.6 r_{\text{cav}}$]

Field size 10 x 10 cm² on surface, SSD = 100 cm: yes no

% $dd(10)_{\text{Pb}}$ (includes e⁻ contamination): $(0.7897/0.9795) \times 100 = 80.62$

50 cm: % $dd(10)_x = [0.8905 + 0.00150\%dd(10)_{\text{Pb}}]\%dd(10)_{\text{Pb}}$ [% $dd(10)_{\text{Pb}} \geq 73\%$] Eq.(13)

30 cm: % $dd(10)_x = [0.8116 + 0.00264\%dd(10)_{\text{Pb}}]\%dd(10)_{\text{Pb}}$ [% $dd(10)_{\text{Pb}} \geq 71\%$] Eq. (14)

If % $dd(10)_{\text{Pb}} < 71\%$ (30 cm) or 73% (50 cm): % $dd(10)_x = \%dd(10)_{\text{Pb}}$

b. % $dd(10)_x$ (for open beam): $[0.8905 + 0.00150(80.62)] [80.62] = 81.54$

Has lead foil been removed? yes no

Interim alternative for energy > 10MV & with ≥ 45 cm clearance: using no lead foil

Measure % $dd(10)$ [% depth-dose at 10 cm depth for curve shifted upstream by $0.6 r_{\text{cav}}$]

% $dd(10)$: No lead $(0.8467/1.055) (100) = 80.26$

% $dd(10)_x = 1.267 (\%dd(10) - 20.0)$ [for 75% $< \%dd(10) \leq 89\%$]

c. % $dd(10)_x =$ $[(1.267) (80.26)] - 20.0 = 81.69$

TG-51 Worksheet A: Photon Beams (cont'd)

5 Determination of k_Q (Sec. IX. B)

Chamber model used to get k_Q :	NEL 2571
a. %dd(10) _x (from 4, above):	81.54
b. k_Q [Table I or Fig. 4]:	0.971

6. Temperature/pressure Correction (Sec.VII.C)

a. Temperature:	20.7	°C
b. Pressure:	750.1 mmHg	kPa $\left[= \text{mmHg} \frac{101.33}{760} \right]$
c. P_{TP} :	1.009	$\left[\text{Eq.(10)} = \left(\frac{273.2 + 6a}{295.2} \right) \left(\frac{101.33}{6b} \right) \right]$

7. Polarity correction (Sec. VII. A.)

M _{raw} ⁺ :	+ 4.061	C or rdg
M _{raw} ⁻ :	- 4.073	C or rdg
a. M _{raw} (for polarity of calibration):	- 4.073	C or rdg
b. P_{pol} :	0.9985	$\left[\text{Eq.(9)} = \left \frac{M_{raw}^+ - M_{raw}^-}{2M_{raw}} \right \right]$

8. P_{ion} measurements (Sec.VII. D. 2)

Operating voltage = V_H :	- 300	V
Lower voltage V_L :	- 150	V
M _{raw} ^H :	- 4.073	C or rdg
M _{raw} ^L :	- 4.0526	C or rdg
⁶⁰ Co treated as general recombination		
a. $P_{ion}(V_H)$ (Eq.(11));		$\left[\left(1 - \left(\frac{V_H}{V_L} \right)^2 \right) / \left(\frac{M_{raw}^H}{M_{raw}^L} - \left(\frac{V_H}{V_L} \right)^2 \right) \right]$
Pulsed/swept beams		
b. $P_{ion}(V_H)$ (Eq.(12))	1.005	$\left[\left(1 - \frac{V_H}{V_L} \right) / \left(\frac{M_{raw}^H}{M_{raw}^L} - \frac{V_H}{V_L} \right) \right]$

If $P_{ion} > 1.05$, another ion chamber should be used.

9. Corrected ion. ch. rdg. M (Sec.VII) at 10 cm depth, water equivalent

$$M = P_{ion} P_{TP} P_{elec} P_{Pol} M_{raw} = [8(a\text{-or-}b) \cdot 6c \cdot 2bi \cdot 7b \cdot 7a]$$

Fully corrected M (Eq.(8)): $\frac{1.005 \times 1.009 \times 0.998 \times 10^{-8} \times 0.9985 \times 4.073}{1} = 4.116 \times 10^{-8} \text{C}$

10. Dose to water at 10 cm depth:

$$D_w^O = M k_Q N_{D,w}^{60Co} = [9 \cdot 5b \cdot 2c] \text{ Eq. (3)}$$

a. Dose to water at 10 cm depth = $\frac{4.116 \times 10^{-8} \times 0.971 \times 4.533 \times 10^7}{1} = 1.8116$

b. Dose/ MU (or min ⁶⁰Co) at 10 cm depth 0.9058 x 10⁻² Gy/MU [10a/3c]

11. Dose to water/MU (or min, ⁶⁰Co) at d_{max} (if relevant locally)

a. Clinical %dd(10) for SSD setup / 100.:
or clinical TMR(10, 10 x 10) for SAD setup:

0.901

b. Dose / MU (or min, ⁶⁰Co) at d_{max} :

$\frac{1.005 \times 10^{-2}}{1.005}$ Gy/MU [10b/(11a)]

1.005 cGy (water)/MU